Maxxam
A Bureau Veritas Group Company

Your Project #: PFC
Site Location: BFTA
Your C.O.C. #: 558437-10-01

Attention:Tom Cambareri

Cape Cod Comission
Cape Cod Commission
3225 Main Street
Barnstable, MA
USA 02630

Report Date: 2016/12/23
Report #: R4299994
Version: 1 - Final

CERTIFICATE OF ANALYSIS

MAXXAM JOB #: B6Q9307
Received: 2016/12/09, 15:15

Sample Matrix: Water
Samples Received: 5

Analyses	Quantity	Date Extracted	Date Analyzed	Laboratory Method	Reference
PFOS and PFOA in water	2	2016/12/13	2016/12/16	CAM SOP-00894	EPA 537 m
PFOS and PFOA in water	3	2016/12/15	2016/12/20	CAM SOP-00894	EPA 537 m

Remarks:

Maxxam Analytics' laboratories are accredited to ISO/IEC 17025:2005 for specific parameters on scopes of accreditation. Unless otherwise noted, procedures used by Maxxam are based upon recognized Provincial, Federal or US method compendia such as CCME, MDDELCC, EPA, APHA.

All work recorded herein has been done in accordance with procedures and practices ordinarily exercised by professionals in Maxxam's profession using accepted testing methodologies, quality assurance and quality control procedures (except where otherwise agreed by the client and Maxxam in writing). All data is in statistical control and has met quality control and method performance criteria unless otherwise noted. All method blanks are reported: unless indicated otherwise, associated sample data are not blank corrected.

Maxxam Analytics' liability is limited to the actual cost of the requested analyses, unless otherwise agreed in writing. There is no other warranty expressed or implied. Maxxam has been retained to provide analysis of samples provided by the Client using the testing methodology referenced in this report. Interpretation and use of test results are the sole responsibility of the Client and are not within the scope of services provided by Maxxam, unless otherwise agreed in writing.

Solid sample results, except biota, are based on dry weight unless otherwise indicated. Organic analyses are not recovery corrected except for isotope dilution methods. Results relate to samples tested.
This Certificate shall not be reproduced except in full, without the written approval of the laboratory.
Reference Method suffix "m" indicates test methods incorporate validated modifications from specific reference methods to improve performance.
* RPDs calculated using raw data. The rounding of final results may result in the apparent difference.

Encryption Key

Stephanie Pollen
Project Manager Assistant
23 Dec 2016 13:35:53

Please direct all questions regarding this Certificate of Analysis to your Project Manager.
Melissa DiGrazia, Project Manager - ATUT
Email: MDiGrazia@maxxam.ca
Phone# (905) 817-5700
==========
Maxxam has procedures in place to guard against improper use of the electronic signature and have the required "signatories", as per section 5.10.2 of ISO/IEC 17025:2005(E), signing the reports. For Service Group specific validation please refer to the Validation Signature Page.

Maxxam Analytics International Corporation o/a Maxxam Analytics 6740 Campobello Road, Mississauga, Ontario, L5N 2L8 Tel: (905) 817-5700 Toll-Free: 800-563-6266 Fax: (905) 817-5777 www.maxxam.ca

Maxxam Job #: B6Q9307
Report Date: 2016/12/23

Cape Cod Comission
Client Project #: PFC
Site Location: BFTA

RESULTS OF ANALYSES OF WATER

Maxxam ID		DPK026	DPK027				DPK028			
Sampling Date		2016/12/08 11:15	2016/12/08 11:30				2016/12/08 12:10			
COC Number		558437-10-01	558437-10-01				558437-10-01			
	UNITS	**PFW-2**	**HSW-1**	**RDL**	**MDL**	**QC Batch**	**FLINT ROCK POND**	**RDL**	**MDL**	**QC Batch**
Miscellaneous Parameters										
6:2 Fluorotelomer sulfonate	ug/L	4.8 (1)	14 (1)	0.80	0.21	4794191	0.020	0.020	0.0065	4794669
8:2 Fluorotelomer sulfonate	ug/L	0.63 (1)	1.3 (1)	0.80	0.28	4794191	<0.0055	0.020	0.0055	4794669
N-ethylperfluorooctane sulfonamide	ug/L	<0.28 (1)	<0.28 (1)	0.80	0.28	4794191	<0.0053	0.020	0.0053	4794669
N-ethylperfluorooctane sulfonamidoe	ug/L	<0.29 (1)	<0.29 (1)	0.80	0.29	4794191	<0.0049	0.020	0.0049	4794669
N-methylperfluorooctane sulfonamide	ug/L	<0.15 (1)	<0.15 (1)	0.80	0.15	4794191	<0.0040	0.020	0.0040	4794669
N-methylperfluorooctanesulfonamidol	ug/L	<0.30 (1)	<0.30 (1)	0.80	0.30	4794191	<0.0061	0.020	0.0061	4794669
Perfluorobutane Sulfonate (PFBS)	ug/L	0.35 (1)	0.57 (1)	0.80	0.23	4794191	0.045	0.020	0.0019	4794669
Perfluorobutanoic acid	ug/L	0.68 (1)	1.0 (1)	0.80	0.20	4794191	0.065	0.020	0.0066	4794669
Perfluorodecane Sulfonate	ug/L	<0.22 (1)	<0.22 (1)	0.80	0.22	4794191	<0.0043	0.020	0.0043	4794669
Perfluorodecanoic Acid (PFDA)	ug/L	<0.20 (1)	<0.20 (1)	0.80	0.20	4794191	0.019	0.020	0.0066	4794669
Perfluorododecanoic Acid (PFDoA)	ug/L	<0.16 (1)	<0.16 (1)	0.80	0.16	4794191	<0.0057	0.020	0.0057	4794669
Perfluoroheptane sulfonate	ug/L	0.40 (1)	0.53 (1)	0.80	0.27	4794191	0.037	0.020	0.0036	4794669
Perfluoroheptanoic Acid (PFHpA)	ug/L	0.52 (1)	1.3 (1)	0.80	0.27	4794191	0.17	0.020	0.0047	4794669
Perfluorohexane Sulfonate (PFHxS)	ug/L	3.2 (1)	7.7 (1)	0.80	0.16	4794191	0.39	0.020	0.0040	4794669
Perfluorohexanoic Acid (PFHxA)	ug/L	3.2 (1)	5.2 (1)	0.80	0.17	4794191	0.35	0.020	0.0046	4794669
Perfluoro-n-Octanoic Acid (PFOA)	ug/L	0.66 (1)	1.8 (1)	0.80	0.20	4794191	0.13	0.020	0.0053	4794669
Perfluorononanoic Acid (PFNA)	ug/L	<0.19 (1)	0.35 (1)	0.80	0.19	4794191	0.11	0.020	0.0046	4794669
Perfluorooctane Sulfonamide (PFOSA)	ug/L	<0.23 (1)	<0.23 (1)	0.80	0.23	4794191	<0.0058	0.020	0.0058	4794669
Perfluorooctane Sulfonate (PFOS)	ug/L	13 (1)	36 (1)	0.80	0.14	4794191	1.3 (2)	0.20	0.033	4794669
Perfluoropentanoic Acid (PFPeA)	ug/L	1.9 (1)	3.0 (1)	0.80	0.21	4794191	0.20	0.020	0.0036	4794669
Perfluorotetradecanoic Acid	ug/L	0.26 (1)	<0.20 (1)	0.80	0.20	4794191	<0.0052	0.020	0.0052	4794669
Perfluorotridecanoic Acid	ug/L	<0.30 (1)	<0.30 (1)	0.80	0.30	4794191	0.0073	0.020	0.0032	4794669
Perfluoroundecanoic Acid (PFUnA)	ug/L	0.48 (1)	0.41 (1)	0.80	0.14	4794191	0.018	0.020	0.0037	4794669
Surrogate Recovery (%)										
13C4-Perfluorooctanesulfonate	%	100	102	N/A	N/A	4794191	91	N/A	N/A	4794669
13C4-Perfluorooctanoic acid	%	104	100	N/A	N/A	4794191	97	N/A	N/A	4794669
13C8-Perfluorooctanesulfonamide	%	94	96	N/A	N/A	4794191	85	N/A	N/A	4794669

RDL = Reportable Detection Limit
QC Batch = Quality Control Batch
N/A = Not Applicable
(1) Due to high concentrations of the target analytes, sample required high level analysis. Detection limit was adjusted accordingly.

(2) Due to high concentration of the target analyte, sample required 10x dilution. Detection limit was adjusted accordingly.

Maxxam Analytics International Corporation o/a Maxxam Analytics 6740 Campobello Road, Mississauga, Ontario, L5N 2L8 Tel: (905) 817-5700 Toll-Free: 800-563-6266 Fax: (905) 817-5777 www.maxxam.ca

Maxxam Job #: B6Q9307
Report Date: 2016/12/23

Cape Cod Comission
Client Project #: PFC
Site Location: BFTA

RESULTS OF ANALYSES OF WATER

Maxxam ID		DPK029	DPK030			
Sampling Date		2016/12/08 10:30	2016/12/08			
COC Number		558437-10-01	558437-10-01			
	UNITS	**RUNOFF AT DIG OUT**	**RUN OFF**	**RDL**	**MDL**	**QC Batch**
Miscellaneous Parameters						
6:2 Fluorotelomer sulfonate	ug/L	0.062	0.0084	0.020	0.0065	4794669
8:2 Fluorotelomer sulfonate	ug/L	0.018	<0.0055	0.020	0.0055	4794669
N-ethylperfluorooctane sulfonamide	ug/L	<0.0053	<0.0053	0.020	0.0053	4794669
N-ethylperfluorooctane sulfonamidoe	ug/L	<0.0049	<0.0049	0.020	0.0049	4794669
N-methylperfluorooctane sulfonamide	ug/L	<0.0040	<0.0040	0.020	0.0040	4794669
N-methylperfluorooctanesulfonamidol	ug/L	<0.0061	<0.0061	0.020	0.0061	4794669
Perfluorobutane Sulfonate (PFBS)	ug/L	<0.0019	<0.0019	0.020	0.0019	4794669
Perfluorobutanoic acid	ug/L	<0.0066	<0.0066	0.020	0.0066	4794669
Perfluorodecane Sulfonate	ug/L	<0.0043	<0.0043	0.020	0.0043	4794669
Perfluorodecanoic Acid (PFDA)	ug/L	<0.0066	<0.0066	0.020	0.0066	4794669
Perfluorododecanoic Acid (PFDoA)	ug/L	<0.0057	<0.0057	0.020	0.0057	4794669
Perfluoroheptane sulfonate	ug/L	<0.0036	<0.0036	0.020	0.0036	4794669
Perfluoroheptanoic Acid (PFHpA)	ug/L	0.0065	<0.0047	0.020	0.0047	4794669
Perfluorohexane Sulfonate (PFHxS)	ug/L	0.15	0.0065	0.020	0.0040	4794669
Perfluorohexanoic Acid (PFHxA)	ug/L	0.0054	<0.0046	0.020	0.0046	4794669
Perfluoro-n-Octanoic Acid (PFOA)	ug/L	0.0066	<0.0053	0.020	0.0053	4794669
Perfluorononanoic Acid (PFNA)	ug/L	<0.0046	<0.0046	0.020	0.0046	4794669
Perfluorooctane Sulfonamide (PFOSA)	ug/L	<0.0058	<0.0058	0.020	0.0058	4794669
Perfluorooctane Sulfonate (PFOS)	ug/L	0.030	0.016	0.020	0.0033	4794669
Perfluoropentanoic Acid (PFPeA)	ug/L	0.0094	0.0085	0.020	0.0036	4794669
Perfluorotetradecanoic Acid	ug/L	<0.0052	<0.0052	0.020	0.0052	4794669
Perfluorotridecanoic Acid	ug/L	0.0067	<0.0032	0.020	0.0032	4794669
Perfluoroundecanoic Acid (PFUnA)	ug/L	0.0063	0.0039	0.020	0.0037	4794669
Surrogate Recovery (%)						
13C4-Perfluorooctanesulfonate	%	81	94	N/A	N/A	4794669
13C4-Perfluorooctanoic acid	%	83	97	N/A	N/A	4794669
13C8-Perfluorooctanesulfonamide	%	86	82	N/A	N/A	4794669

RDL = Reportable Detection Limit
QC Batch = Quality Control Batch
N/A = Not Applicable

Maxxam Analytics International Corporation o/a Maxxam Analytics 6740 Campobello Road, Mississauga, Ontario, L5N 2L8 Tel: (905) 817-5700 Toll-Free: 800-563-6266 Fax: (905) 817-5777 www.maxxam.ca

Maxxam Job #: B6Q9307
Report Date: 2016/12/23

Cape Cod Comission
Client Project #: PFC
Site Location: BFTA

TEST SUMMARY

Maxxam ID: DPK026 **Collected:** 2016/12/08
Sample ID: PFW-2 **Shipped:**
Matrix: Water **Received:** 2016/12/09

Test Description	Instrumentation	Batch	Extracted	Date Analyzed	Analyst
PFOS and PFOA in water	LCMS	4794191	2016/12/13	2016/12/16	Colm McNamara

Maxxam ID: DPK027 **Collected:** 2016/12/08
Sample ID: HSW-1 **Shipped:**
Matrix: Water **Received:** 2016/12/09

Test Description	Instrumentation	Batch	Extracted	Date Analyzed	Analyst
PFOS and PFOA in water	LCMS	4794191	2016/12/13	2016/12/16	Colm McNamara

Maxxam ID: DPK028 **Collected:** 2016/12/08
Sample ID: FLINT ROCK POND **Shipped:**
Matrix: Water **Received:** 2016/12/09

Test Description	Instrumentation	Batch	Extracted	Date Analyzed	Analyst
PFOS and PFOA in water	LCMS	4794669	2016/12/15	2016/12/20	Colm McNamara

Maxxam ID: DPK029 **Collected:** 2016/12/08
Sample ID: RUNOFF AT DIG OUT **Shipped:**
Matrix: Water **Received:** 2016/12/09

Test Description	Instrumentation	Batch	Extracted	Date Analyzed	Analyst
PFOS and PFOA in water	LCMS	4794669	2016/12/15	2016/12/20	Colm McNamara

Maxxam ID: DPK030 **Collected:** 2016/12/08
Sample ID: RUN OFF **Shipped:**
Matrix: Water **Received:** 2016/12/09

Test Description	Instrumentation	Batch	Extracted	Date Analyzed	Analyst
PFOS and PFOA in water	LCMS	4794669	2016/12/15	2016/12/20	Colm McNamara

Maxxam Analytics International Corporation o/a Maxxam Analytics 6740 Campobello Road, Mississauga, Ontario, L5N 2L8 Tel: (905) 817-5700 Toll-Free: 800-563-6266 Fax: (905) 817-5777 www.maxxam.ca

GENERAL COMMENTS

Results relate only to the items tested.

Maxxam Analytics International Corporation o/a Maxxam Analytics 6740 Campobello Road, Mississauga, Ontario, L5N 2L8 Tel: (905) 817-5700 Toll-Free: 800-563-6266 Fax: (905) 817-5777 www.maxxam.ca

Maxxam Job #: B6Q9307
Report Date: 2016/12/23

Cape Cod Comission
Client Project #: PFC
Site Location: BFTA

QUALITY ASSURANCE REPORT

QA/QC Batch	Init	QC Type	Parameter	Date Analyzed	Value	% Recovery	UNITS	QC Limits
4794191	CM5	Matrix Spike(DPK027)	13C4-Perfluorooctanesulfonate	2016/12/16		91	%	70 - 130
			13C4-Perfluorooctanoic acid	2016/12/16		88	%	70 - 130
			13C8-Perfluorooctanesulfonamide	2016/12/16		87	%	60 - 120
			6:2 Fluorotelomer sulfonate	2016/12/16		96	%	70 - 130
			8:2 Fluorotelomer sulfonate	2016/12/16		99	%	70 - 130
			N-ethylperfluorooctane sulfonamide	2016/12/16		98	%	70 - 130
			N-ethylperfluorooctane sulfonamidoe	2016/12/16		90	%	70 - 130
			N-methylperfluorooctane sulfonamide	2016/12/16		100	%	70 - 130
			N-methylperfluorooctanesulfonamidol	2016/12/16		88	%	70 - 130
			Perfluorobutane Sulfonate (PFBS)	2016/12/16		91	%	70 - 130
			Perfluorobutanoic acid	2016/12/16		112	%	70 - 130
			Perfluorodecane Sulfonate	2016/12/16		111	%	70 - 130
			Perfluoroheptane sulfonate	2016/12/16		92	%	70 - 130
			Perfluoroheptanoic Acid (PFHpA)	2016/12/16		100	%	70 - 130
			Perfluorohexane Sulfonate (PFHxS)	2016/12/16		99	%	70 - 130
			Perfluorohexanoic Acid (PFHxA)	2016/12/16		97	%	70 - 130
			Perfluorononanoic Acid (PFNA)	2016/12/16		103	%	70 - 130
			Perfluorooctane Sulfonamide (PFOSA)	2016/12/16		101	%	70 - 130
			Perfluoropentanoic Acid (PFPeA)	2016/12/16		102	%	70 - 130
			Perfluorotetradecanoic Acid	2016/12/16		110	%	70 - 130
			Perfluorotridecanoic Acid	2016/12/16		105	%	70 - 130
			Perfluoroundecanoic Acid (PFUnA)	2016/12/16		99	%	70 - 130
			Perfluorodecanoic Acid (PFDA)	2016/12/16		105	%	70 - 130
			Perfluorododecanoic Acid (PFDoA)	2016/12/16		104	%	70 - 130
			Perfluoro-n-Octanoic Acid (PFOA)	2016/12/16		101	%	70 - 130
			Perfluorooctane Sulfonate (PFOS)	2016/12/16		NC	%	70 - 130
4794191	CM5	Matrix Spike DUP(DPK027)	13C4-Perfluorooctanesulfonate	2016/12/16		100	%	70 - 130
			13C4-Perfluorooctanoic acid	2016/12/16		99	%	70 - 130
			13C8-Perfluorooctanesulfonamide	2016/12/16		83	%	60 - 120
			6:2 Fluorotelomer sulfonate	2016/12/16		100	%	70 - 130
			8:2 Fluorotelomer sulfonate	2016/12/16		107	%	70 - 130
			N-ethylperfluorooctane sulfonamide	2016/12/16		95	%	70 - 130
			N-ethylperfluorooctane sulfonamidoe	2016/12/16		105	%	70 - 130
			N-methylperfluorooctane sulfonamide	2016/12/16		104	%	70 - 130
			N-methylperfluorooctanesulfonamidol	2016/12/16		86	%	70 - 130
			Perfluorobutane Sulfonate (PFBS)	2016/12/16		102	%	70 - 130
			Perfluorobutanoic acid	2016/12/16		114	%	70 - 130
			Perfluorodecane Sulfonate	2016/12/16		104	%	70 - 130
			Perfluoroheptane sulfonate	2016/12/16		103	%	70 - 130
			Perfluoroheptanoic Acid (PFHpA)	2016/12/16		102	%	70 - 130
			Perfluorohexane Sulfonate (PFHxS)	2016/12/16		105	%	70 - 130
			Perfluorohexanoic Acid (PFHxA)	2016/12/16		102	%	70 - 130
			Perfluorononanoic Acid (PFNA)	2016/12/16		109	%	70 - 130
			Perfluorooctane Sulfonamide (PFOSA)	2016/12/16		115	%	70 - 130
			Perfluoropentanoic Acid (PFPeA)	2016/12/16		102	%	70 - 130
			Perfluorotetradecanoic Acid	2016/12/16		118	%	70 - 130
			Perfluorotridecanoic Acid	2016/12/16		109	%	70 - 130
			Perfluoroundecanoic Acid (PFUnA)	2016/12/16		104	%	70 - 130
			Perfluorodecanoic Acid (PFDA)	2016/12/16		110	%	70 - 130
			Perfluorododecanoic Acid (PFDoA)	2016/12/16		99	%	70 - 130
			Perfluoro-n-Octanoic Acid (PFOA)	2016/12/16		104	%	70 - 130
			Perfluorooctane Sulfonate (PFOS)	2016/12/16		NC	%	70 - 130
4794191	CM5	MS/MSD RPD	6:2 Fluorotelomer sulfonate	2016/12/16	4.7		%	30

Maxxam Analytics International Corporation o/a Maxxam Analytics 6740 Campobello Road, Mississauga, Ontario, L5N 2L8 Tel: (905) 817-5700 Toll-Free: 800-563-6266 Fax: (905) 817-5777 www.maxxam.ca

Maxxam Job #: B6Q9307
Report Date: 2016/12/23

Cape Cod Comission
Client Project #: PFC
Site Location: BFTA

QUALITY ASSURANCE REPORT(CONT'D)

QA/QC Batch	Init	QC Type	Parameter	Date Analyzed	Value	% Recovery	UNITS	QC Limits
			8:2 Fluorotelomer sulfonate	2016/12/16	7.6		%	30
			N-ethylperfluorooctane sulfonamide	2016/12/16	2.6		%	30
			N-ethylperfluorooctane sulfonamidoe	2016/12/16	15		%	30
			N-methylperfluorooctane sulfonamide	2016/12/16	3.8		%	30
			N-methylperfluorooctanesulfonamidol	2016/12/16	1.6		%	30
			Perfluorobutane Sulfonate (PFBS)	2016/12/16	12		%	30
			Perfluorobutanoic acid	2016/12/16	1.3		%	30
			Perfluorodecane Sulfonate	2016/12/16	6.3		%	30
			Perfluoroheptane sulfonate	2016/12/16	11		%	30
			Perfluoroheptanoic Acid (PFHpA)	2016/12/16	1.8		%	30
			Perfluorohexane Sulfonate (PFHxS)	2016/12/16	6.0		%	30
			Perfluorohexanoic Acid (PFHxA)	2016/12/16	4.3		%	30
			Perfluorononanoic Acid (PFNA)	2016/12/16	5.1		%	30
			Perfluorooctane Sulfonamide (PFOSA)	2016/12/16	13		%	30
			Perfluoropentanoic Acid (PFPeA)	2016/12/16	0.35		%	30
			Perfluorotetradecanoic Acid	2016/12/16	7.5		%	30
			Perfluorotridecanoic Acid	2016/12/16	3.0		%	30
			Perfluoroundecanoic Acid (PFUnA)	2016/12/16	4.2		%	30
			Perfluorodecanoic Acid (PFDA)	2016/12/16	4.6		%	30
			Perfluorododecanoic Acid (PFDoA)	2016/12/16	4.9		%	30
			Perfluoro-n-Octanoic Acid (PFOA)	2016/12/16	2.8		%	30
			Perfluorooctane Sulfonate (PFOS)	2016/12/16	NC		%	30
4794191	CM5	Spiked Blank	13C4-Perfluorooctanesulfonate	2016/12/16		97	%	70 - 130
			13C4-Perfluorooctanoic acid	2016/12/16		98	%	70 - 130
			13C8-Perfluorooctanesulfonamide	2016/12/16		97	%	60 - 120
			6:2 Fluorotelomer sulfonate	2016/12/16		104	%	70 - 130
			8:2 Fluorotelomer sulfonate	2016/12/16		106	%	70 - 130
			N-ethylperfluorooctane sulfonamide	2016/12/16		104	%	70 - 130
			N-ethylperfluorooctane sulfonamidoe	2016/12/16		102	%	70 - 130
			N-methylperfluorooctane sulfonamide	2016/12/16		100	%	70 - 130
			N-methylperfluorooctanesulfonamidol	2016/12/16		97	%	70 - 130
			Perfluorobutane Sulfonate (PFBS)	2016/12/16		111	%	70 - 130
			Perfluorobutanoic acid	2016/12/16		105	%	70 - 130
			Perfluorodecane Sulfonate	2016/12/16		104	%	70 - 130
			Perfluoroheptane sulfonate	2016/12/16		99	%	70 - 130
			Perfluoroheptanoic Acid (PFHpA)	2016/12/16		101	%	70 - 130
			Perfluorohexane Sulfonate (PFHxS)	2016/12/16		110	%	70 - 130
			Perfluorohexanoic Acid (PFHxA)	2016/12/16		101	%	70 - 130
			Perfluorononanoic Acid (PFNA)	2016/12/16		110	%	70 - 130
			Perfluorooctane Sulfonamide (PFOSA)	2016/12/16		109	%	70 - 130
			Perfluoropentanoic Acid (PFPeA)	2016/12/16		106	%	70 - 130
			Perfluorotetradecanoic Acid	2016/12/16		109	%	70 - 130
			Perfluorotridecanoic Acid	2016/12/16		111	%	70 - 130
			Perfluoroundecanoic Acid (PFUnA)	2016/12/16		105	%	70 - 130
			Perfluorodecanoic Acid (PFDA)	2016/12/16		109	%	70 - 130
			Perfluorododecanoic Acid (PFDoA)	2016/12/16		113	%	70 - 130
			Perfluoro-n-Octanoic Acid (PFOA)	2016/12/16		108	%	70 - 130
			Perfluorooctane Sulfonate (PFOS)	2016/12/16		108	%	70 - 130
4794191	CM5	Method Blank	13C4-Perfluorooctanesulfonate	2016/12/16		106	%	70 - 130
			13C4-Perfluorooctanoic acid	2016/12/16		96	%	70 - 130
			13C8-Perfluorooctanesulfonamide	2016/12/16		89	%	60 - 120
			6:2 Fluorotelomer sulfonate	2016/12/16	<0.21		ug/L	
			8:2 Fluorotelomer sulfonate	2016/12/16	<0.28		ug/L	

Maxxam Analytics International Corporation o/a Maxxam Analytics 6740 Campobello Road, Mississauga, Ontario, L5N 2L8 Tel: (905) 817-5700 Toll-Free: 800-563-6266 Fax: (905) 817-5777 www.maxxam.ca

Maxxam Job #: B6Q9307
Report Date: 2016/12/23

Cape Cod Comission
Client Project #: PFC
Site Location: BFTA

QUALITY ASSURANCE REPORT(CONT'D)

QA/QC Batch	Init	QC Type	Parameter	Date Analyzed	Value	% Recovery	UNITS	QC Limits
			N-ethylperfluorooctane sulfonamide	2016/12/16	<0.28		ug/L	
			N-ethylperfluorooctane sulfonamidoe	2016/12/16	<0.29		ug/L	
			N-methylperfluorooctane sulfonamide	2016/12/16	<0.15		ug/L	
			N-methylperfluorooctanesulfonamidol	2016/12/16	<0.30		ug/L	
			Perfluorobutane Sulfonate (PFBS)	2016/12/16	<0.23		ug/L	
			Perfluorobutanoic acid	2016/12/16	<0.20		ug/L	
			Perfluorodecane Sulfonate	2016/12/16	<0.22		ug/L	
			Perfluoroheptane sulfonate	2016/12/16	<0.27		ug/L	
			Perfluoroheptanoic Acid (PFHpA)	2016/12/16	<0.27		ug/L	
			Perfluorohexane Sulfonate (PFHxS)	2016/12/16	<0.16		ug/L	
			Perfluorohexanoic Acid (PFHxA)	2016/12/16	<0.17		ug/L	
			Perfluorononanoic Acid (PFNA)	2016/12/16	<0.19		ug/L	
			Perfluorooctane Sulfonamide (PFOSA)	2016/12/16	<0.23		ug/L	
			Perfluoropentanoic Acid (PFPeA)	2016/12/16	<0.21		ug/L	
			Perfluorotetradecanoic Acid	2016/12/16	<0.20		ug/L	
			Perfluorotridecanoic Acid	2016/12/16	<0.30		ug/L	
			Perfluoroundecanoic Acid (PFUnA)	2016/12/16	<0.14		ug/L	
			Perfluorodecanoic Acid (PFDA)	2016/12/16	<0.20		ug/L	
			Perfluorododecanoic Acid (PFDoA)	2016/12/16	<0.16		ug/L	
			Perfluoro-n-Octanoic Acid (PFOA)	2016/12/16	<0.20		ug/L	
			Perfluorooctane Sulfonate (PFOS)	2016/12/16	<0.14		ug/L	
4794669	CM5	Spiked Blank	13C4-Perfluorooctanesulfonate	2016/12/20		79	%	70 - 130
			13C4-Perfluorooctanoic acid	2016/12/20		88	%	70 - 130
			13C8-Perfluorooctanesulfonamide	2016/12/20		74	%	60 - 120
			6:2 Fluorotelomer sulfonate	2016/12/20		106	%	70 - 130
			8:2 Fluorotelomer sulfonate	2016/12/20		110	%	70 - 130
			N-ethylperfluorooctane sulfonamide	2016/12/20		108	%	70 - 130
			N-ethylperfluorooctane sulfonamidoe	2016/12/20		93	%	70 - 130
			N-methylperfluorooctane sulfonamide	2016/12/20		106	%	70 - 130
			N-methylperfluorooctanesulfonamidol	2016/12/20		119	%	70 - 130
			Perfluorobutane Sulfonate (PFBS)	2016/12/20		115	%	70 - 130
			Perfluorobutanoic acid	2016/12/20		113	%	70 - 130
			Perfluorodecane Sulfonate	2016/12/20		92	%	70 - 130
			Perfluoroheptane sulfonate	2016/12/20		102	%	70 - 130
			Perfluoroheptanoic Acid (PFHpA)	2016/12/20		117	%	70 - 130
			Perfluorohexane Sulfonate (PFHxS)	2016/12/20		113	%	70 - 130
			Perfluorohexanoic Acid (PFHxA)	2016/12/20		122	%	70 - 130
			Perfluorononanoic Acid (PFNA)	2016/12/20		116	%	70 - 130
			Perfluorooctane Sulfonamide (PFOSA)	2016/12/20		117	%	70 - 130
			Perfluoropentanoic Acid (PFPeA)	2016/12/20		112	%	70 - 130
			Perfluorotetradecanoic Acid	2016/12/20		113	%	70 - 130
			Perfluorotridecanoic Acid	2016/12/20		116	%	70 - 130
			Perfluoroundecanoic Acid (PFUnA)	2016/12/20		120	%	70 - 130
			Perfluorodecanoic Acid (PFDA)	2016/12/20		119	%	70 - 130
			Perfluorododecanoic Acid (PFDoA)	2016/12/20		115	%	70 - 130
			Perfluoro-n-Octanoic Acid (PFOA)	2016/12/20		111	%	70 - 130
			Perfluorooctane Sulfonate (PFOS)	2016/12/20		117	%	70 - 130
4794669	CM5	Spiked Blank DUP	13C4-Perfluorooctanesulfonate	2016/12/20		84	%	70 - 130
			13C4-Perfluorooctanoic acid	2016/12/20		89	%	70 - 130
			13C8-Perfluorooctanesulfonamide	2016/12/20		82	%	60 - 120
			6:2 Fluorotelomer sulfonate	2016/12/20		113	%	70 - 130
			8:2 Fluorotelomer sulfonate	2016/12/20		123	%	70 - 130
			N-ethylperfluorooctane sulfonamide	2016/12/20		96	%	70 - 130

Maxxam Analytics International Corporation o/a Maxxam Analytics 6740 Campobello Road, Mississauga, Ontario, L5N 2L8 Tel: (905) 817-5700 Toll-Free: 800-563-6266 Fax: (905) 817-5777 www.maxxam.ca

Maxxam Job #: B6Q9307
Report Date: 2016/12/23

Cape Cod Comission
Client Project #: PFC
Site Location: BFTA

QUALITY ASSURANCE REPORT(CONT'D)

QA/QC Batch	Init	QC Type	Parameter	Date Analyzed	Value	% Recovery	UNITS	QC Limits
			N-ethylperfluorooctane sulfonamidoe	2016/12/20		98	%	70 - 130
			N-methylperfluorooctane sulfonamide	2016/12/20		102	%	70 - 130
			N-methylperfluorooctanesulfonamidol	2016/12/20		114	%	70 - 130
			Perfluorobutane Sulfonate (PFBS)	2016/12/20		111	%	70 - 130
			Perfluorobutanoic acid	2016/12/20		112	%	70 - 130
			Perfluorodecane Sulfonate	2016/12/20		107	%	70 - 130
			Perfluoroheptane sulfonate	2016/12/20		106	%	70 - 130
			Perfluoroheptanoic Acid (PFHpA)	2016/12/20		119	%	70 - 130
			Perfluorohexane Sulfonate (PFHxS)	2016/12/20		113	%	70 - 130
			Perfluorohexanoic Acid (PFHxA)	2016/12/20		125	%	70 - 130
			Perfluorononanoic Acid (PFNA)	2016/12/20		119	%	70 - 130
			Perfluorooctane Sulfonamide (PFOSA)	2016/12/20		116	%	70 - 130
			Perfluoropentanoic Acid (PFPeA)	2016/12/20		110	%	70 - 130
			Perfluorotetradecanoic Acid	2016/12/20		113	%	70 - 130
			Perfluorotridecanoic Acid	2016/12/20		115	%	70 - 130
			Perfluoroundecanoic Acid (PFUnA)	2016/12/20		115	%	70 - 130
			Perfluorodecanoic Acid (PFDA)	2016/12/20		111	%	70 - 130
			Perfluorododecanoic Acid (PFDoA)	2016/12/20		114	%	70 - 130
			Perfluoro-n-Octanoic Acid (PFOA)	2016/12/20		111	%	70 - 130
			Perfluorooctane Sulfonate (PFOS)	2016/12/20		110	%	70 - 130
4794669	CM5	RPD	6:2 Fluorotelomer sulfonate	2016/12/20	6.2		%	30
			8:2 Fluorotelomer sulfonate	2016/12/20	11		%	30
			N-ethylperfluorooctane sulfonamide	2016/12/20	11		%	30
			N-ethylperfluorooctane sulfonamidoe	2016/12/20	6.1		%	30
			N-methylperfluorooctane sulfonamide	2016/12/20	4.0		%	30
			N-methylperfluorooctanesulfonamidol	2016/12/20	4.3		%	30
			Perfluorobutane Sulfonate (PFBS)	2016/12/20	2.8		%	30
			Perfluorobutanoic acid	2016/12/20	1.4		%	30
			Perfluorodecane Sulfonate	2016/12/20	15		%	30
			Perfluoroheptane sulfonate	2016/12/20	3.7		%	30
			Perfluoroheptanoic Acid (PFHpA)	2016/12/20	1.7		%	30
			Perfluorohexane Sulfonate (PFHxS)	2016/12/20	0.18		%	30
			Perfluorohexanoic Acid (PFHxA)	2016/12/20	2.6		%	30
			Perfluorononanoic Acid (PFNA)	2016/12/20	2.2		%	30
			Perfluorooctane Sulfonamide (PFOSA)	2016/12/20	0.69		%	30
			Perfluoropentanoic Acid (PFPeA)	2016/12/20	1.8		%	30
			Perfluorotetradecanoic Acid	2016/12/20	0.35		%	30
			Perfluorotridecanoic Acid	2016/12/20	0.86		%	30
			Perfluoroundecanoic Acid (PFUnA)	2016/12/20	3.9		%	30
			Perfluorodecanoic Acid (PFDA)	2016/12/20	6.8		%	30
			Perfluorododecanoic Acid (PFDoA)	2016/12/20	0.88		%	30
			Perfluoro-n-Octanoic Acid (PFOA)	2016/12/20	0.18		%	30
			Perfluorooctane Sulfonate (PFOS)	2016/12/20	6.7		%	30
4794669	CM5	Method Blank	13C4-Perfluorooctanesulfonate	2016/12/20		106	%	70 - 130
			13C4-Perfluorooctanoic acid	2016/12/20		105	%	70 - 130
			13C8-Perfluorooctanesulfonamide	2016/12/20		99	%	60 - 120
			6:2 Fluorotelomer sulfonate	2016/12/20	<0.0065		ug/L	
			8:2 Fluorotelomer sulfonate	2016/12/20	<0.0055		ug/L	
			N-ethylperfluorooctane sulfonamide	2016/12/20	<0.0053		ug/L	
			N-ethylperfluorooctane sulfonamidoe	2016/12/20	<0.0049		ug/L	
			N-methylperfluorooctane sulfonamide	2016/12/20	<0.0040		ug/L	
			N-methylperfluorooctanesulfonamidol	2016/12/20	<0.0061		ug/L	
			Perfluorobutane Sulfonate (PFBS)	2016/12/20	<0.0019		ug/L	

Maxxam Analytics International Corporation o/a Maxxam Analytics 6740 Campobello Road, Mississauga, Ontario, L5N 2L8 Tel: (905) 817-5700 Toll-Free: 800-563-6266 Fax: (905) 817-5777 www.maxxam.ca

Maxxam Job #: B6Q9307
Report Date: 2016/12/23

Cape Cod Comission
Client Project #: PFC
Site Location: BFTA

QUALITY ASSURANCE REPORT(CONT'D)

QA/QC Batch	Init	QC Type	Parameter	Date Analyzed	Value	% Recovery	UNITS	QC Limits
			Perfluorobutanoic acid	2016/12/20	<0.0066		ug/L	
			Perfluorodecane Sulfonate	2016/12/20	<0.0043		ug/L	
			Perfluoroheptane sulfonate	2016/12/20	<0.0036		ug/L	
			Perfluoroheptanoic Acid (PFHpA)	2016/12/20	<0.0047		ug/L	
			Perfluorohexane Sulfonate (PFHxS)	2016/12/20	<0.0040		ug/L	
			Perfluorohexanoic Acid (PFHxA)	2016/12/20	<0.0046		ug/L	
			Perfluorononanoic Acid (PFNA)	2016/12/20	<0.0046		ug/L	
			Perfluorooctane Sulfonamide (PFOSA)	2016/12/20	<0.0058		ug/L	
			Perfluoropentanoic Acid (PFPeA)	2016/12/20	<0.0036		ug/L	
			Perfluorotetradecanoic Acid	2016/12/20	<0.0052		ug/L	
			Perfluorotridecanoic Acid	2016/12/20	<0.0032		ug/L	
			Perfluoroundecanoic Acid (PFUnA)	2016/12/20	<0.0037		ug/L	
			Perfluorodecanoic Acid (PFDA)	2016/12/20	<0.0066		ug/L	
			Perfluorododecanoic Acid (PFDoA)	2016/12/20	<0.0057		ug/L	
			Perfluoro-n-Octanoic Acid (PFOA)	2016/12/20	<0.0053		ug/L	
			Perfluorooctane Sulfonate (PFOS)	2016/12/20	<0.0033		ug/L	

Duplicate: Paired analysis of a separate portion of the same sample. Used to evaluate the variance in the measurement.

Matrix Spike: A sample to which a known amount of the analyte of interest has been added. Used to evaluate sample matrix interference.

Spiked Blank: A blank matrix sample to which a known amount of the analyte, usually from a second source, has been added. Used to evaluate method accuracy.

Method Blank: A blank matrix containing all reagents used in the analytical procedure. Used to identify laboratory contamination.

Surrogate: A pure or isotopically labeled compound whose behavior mirrors the analytes of interest. Used to evaluate extraction efficiency.

NC (Matrix Spike): The recovery in the matrix spike was not calculated. The relative difference between the concentration in the parent sample and the spiked amount was too small to permit a reliable recovery calculation (matrix spike concentration was less than 2x that of the native sample concentration).

Maxxam Analytics International Corporation o/a Maxxam Analytics 6740 Campobello Road, Mississauga, Ontario, L5N 2L8 Tel: (905) 817-5700 Toll-Free: 800-563-6266 Fax: (905) 817-5777 www.maxxam.ca

Maxxam Job #: B6Q9307
Report Date: 2016/12/23

Cape Cod Comission
Client Project #: PFC
Site Location: BFTA

VALIDATION SIGNATURE PAGE

The analytical data and all QC contained in this report were reviewed and validated by the following individual(s).

Adam Robinson, Supervisor, LC/MS/MS

Sin Chii Chia, Scientific Services

Maxxam has procedures in place to guard against improper use of the electronic signature and have the required "signatories", as per section 5.10.2 of ISO/IEC 17025:2005(E), signing the reports. For Service Group specific validation please refer to the Validation Signature Page.

Maxxam Analytics International Corporation o/a Maxxam Analytics 6740 Campobello Road, Mississauga, Ontario, L5N 2L8 Tel: (905) 817-5700 Toll-Free: 800-563-6266 Fax: (905) 817-5777 www.maxxam.ca